

# TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

**Certificate No TSP-06022022-003**

<p><b>Gage ID</b> PRO-4229  <b>Gage S/N</b> 4229  <b>Description</b> 50-0-50 Gauss Meter  <b>Operating Procedure:</b> OP-117  <b>Unit of Meas.</b> Gauss  <b>Manufacturer</b> R. B. Annis Co., Inc.  <b>Cal. Date</b> 6/2/2022  <b>Next Due</b> 12/2/2022  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved Yes</b>  <b>Customer Info.</b> Pro Mag</p>
---	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

SO 10642 JOB 11516

<b>Standard ID</b>	+25	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	24	<b>Nominal</b>	25	<b>Maximum</b>	26
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	25	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	25	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-09-1234						
<b>Std Due Date</b>	5/16/2023	<b>Gage S/N</b>	09-1234	<b>NIST No.</b>	88411		

<b>Standard ID</b>	+50	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	49	<b>Nominal</b>	50	<b>Maximum</b>	51
<b>Ref Type</b>	Gauss Meter	<b>As Found</b>	50	<b>Accuracy</b>	0	<b>Fail Before</b>	No
		<b>As Left</b>	50	<b>Accuracy</b>	0	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-09-1234						
<b>Std Due Date</b>	5/16/2023	<b>Gage S/N</b>	09-1234	<b>NIST No.</b>	88411		

<b>Standard ID</b>	-25	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	24	<b>Nominal</b>	25	<b>Maximum</b>	26

Ref Type	Gauss Meter	As Found	25	Accuracy	0	Fail Before	No
		As Left	25	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date 5/16/2023

Gage S/N 09-1234 NIST No. 88411

Standard ID	-50	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	49	Nominal	50	Maximum	51
Ref Type	Gauss Meter	As Found	50	Accuracy	0	Fail Before	No
		As Left	50	Accuracy	0	Fail After	No

Gage ID of Standard REF-09-1234

Std Due Date 5/16/2023

Gage S/N 09-1234 NIST No. 88411

Calibrated By Taichi Daimo Signature  Date: 6/2/2022