

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-12062018-014

<p>Gage ID PRO-14195 Gage S/N 14195 Description Scope Operating Procedure: SDM-12 Rev.1 Unit of Meas. mm Manufacturer King Cal. Date 12/6/2018 Next Due 12/6/2019 Cal. Freq. 1.00 Years Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Pro Mag</p>
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 17025 and thus 9001-2008 and API-Q1. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 6855 JOB 7457

Gage: 0-6mm

Scope Reading As Found

Master	Actual	Deviation
+0.000000	+0.000000	+0.000000
+1.000000	+1.000000	+0.000000
+2.000000	+2.000000	+0.000000
+3.000000	+3.000000	+0.000000
+4.000000	+4.000000	+0.000000
+5.000000	+5.000000	+0.000000
+6.000000	+6.000000	+0.000000

Tol +0.100000 / -0.100000

Scope Reading As Left

Master	Actual	Deviation
+0.000000	+0.000000	+0.000000
+1.000000	+1.000000	+0.000000
+2.000000	+2.000000	+0.000000
+3.000000	+3.000000	+0.000000
+4.000000	+4.000000	+0.000000

+5.000000 +5.000000 +0.000000
+6.000000 +6.000000 +0.000000

Tol +0.100000 / -0.100000

Procedure # SDM-12 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial #: 17656
Stage Micrometer, Detroit Testing, Model N/A
Traceable to NIST via NAMAS certification no: Cert# CI239-31347-329

Last / Next Cal Dates: 08/27/2018 --> 02/27/2020

Calibrated By Taichi Daimo Signature  Date: 12/6/2018