

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06022021-009

<p>Gage ID PRO-AM26DG0149 Gage S/N AM26DG0149 Description Lead Gauge Operating Procedure: OP-124 Unit of Meas. Inch Manufacturer Gagemaker Cal. Date 6/2/2021 Next Due 12/2/2021 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Pro Mag</p>
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 9635 JOB 10447

Standard ID	+0.00500"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.00450	Nominal	0.00500	Maximum	0.00550
Ref Type	Mic Trac	As Found	0.00504	Accuracy	0.00004	Fail Before	No
		As Left	0.00504	Accuracy	0.00004	Fail After	No

Gage ID of Standard REF-G141308
Std Due Date 9/24/2021 **Gage S/N** G141308 **NIST No.** 747V66-180511-00

Standard ID	+0.01000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.00950	Nominal	0.01000	Maximum	0.01050
Ref Type	Mic Trac	As Found	0.01000	Accuracy	0.00000	Fail Before	No
		As Left	0.01000	Accuracy	0.00000	Fail After	No

Gage ID of Standard REF-G141308
Std Due Date 9/24/2021 **Gage S/N** G141308 **NIST No.** 747V66-180511-00

Standard ID	-0.00500"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	-0.00550	Nominal	-0.00500	Maximum	-0.00450

