

# TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

Certificate No TSP-09042020-006

<p><b>Gage ID</b> PRO-E12933  <b>Gage S/N</b> E12933  <b>Description</b> Pit Gauge  <b>Operating Procedure:</b> OP-122  <b>Unit of Meas.</b> Inch  <b>Manufacturer</b> Gagemaker  <b>Cal. Date</b> 9/4/2020  <b>Next Due</b> 3/4/2021  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Pro Mag</p>
---	---

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

SO 8935 JOB 9660

<b>Standard ID</b> .10000 "	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 0.09900	<b>Nominal</b> 0.10000	<b>Maximum</b> 0.10100
<b>Ref Type</b> Mic Trac	<b>As Found</b> 0.10005	<b>Accuracy</b> 0.00005	<b>Fall Before</b> No
	<b>As Left</b> 0.10005	<b>Accuracy</b> 0.00005	<b>Fall After</b> No

**Gage ID of Standard** REF-G141308  
**Std Due Date** 9/18/2020      **Gage S/N** G141308      **NIST No.** 747V66-180511-00

<b>Standard ID</b> .20000 "	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 0.19900	<b>Nominal</b> 0.20000	<b>Maximum</b> 0.20100
<b>Ref Type</b> Mic Trac	<b>As Found</b> 0.20002	<b>Accuracy</b> 0.00002	<b>Fall Before</b> No
	<b>As Left</b> 0.20002	<b>Accuracy</b> 0.00002	<b>Fall After</b> No

**Gage ID of Standard** REF-G141308  
**Std Due Date** 9/18/2020      **Gage S/N** G141308      **NIST No.** 747V66-180511-00

<b>Standard ID</b> .30000 "	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 0.29900	<b>Nominal</b> 0.30000	<b>Maximum</b> 0.30100

Gage ID: PRO-E12933

Ref Type	Mic Trac	As Found	0.29999	Accuracy	-0.00001	Fail Before	No
		As Left	0.29999	Accuracy	-0.00001	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 9/18/2020

Gage S/N G141308 NIST No. 747V66-180511-00

Standard ID	.40000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.39900	Nominal	0.40000	Maximum	0.40100
Ref Type	Mic Trac	As Found	0.40007	Accuracy	0.00007	Fail Before	No
		As Left	0.40007	Accuracy	0.00007	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 9/18/2020

Gage S/N G141308 NIST No. 747V66-180511-00

Standard ID	.50000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.49900	Nominal	0.50000	Maximum	0.50100
Ref Type	Mic Trac	As Found	0.50015	Accuracy	0.00015	Fail Before	No
		As Left	0.50015	Accuracy	0.00015	Fail After	No

Gage ID of Standard REF-G141308  
 Std Due Date 9/18/2020

Gage S/N G141308 NIST No. 747V66-180511-00

Calibrated By Taichi Daimo Signature  Date: 9/4/2020

9/4/2020

9/4/2020