

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-10252017-008

<p>Gage ID PRO-130103225 Gage S/N 130103225 Description Visible Light Meter Operating Procedure: OP-130 Unit of Meas. Foot Candles Manufacturer Magnaflux Cal. Date 10/25/2017 Next Due 4/25/2018 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Pro Mag</p>
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 17025 and thus 9001-2008 and API-Q1. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 5698 JOB 6280

Using calibrated light meter serial no. 1705427 with sensor 1705430 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	97.3	96.1
Low	33.8	32.9

Readings As Left (foot candles)

	Master	Actual
High	97.3	96.1
Low	33.8	32.9

Cal due date for light meter is 7/14/18
 NIST Traceable through : 166898, 166899, 166900

Calibrated By Taichi Daimo Signature  Date: 10/25/2017