

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-10162018-007

<p>Gage ID PRO-39459 Gage S/N 39459 Description UT Thickness Gage Operating Procedure: OP-121 Unit of Meas. Inch Manufacturer Dakota Ultrasonics Cal. Date 10/16/2018 Next Due 4/16/2019 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Pro Mag</p>
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 17025 and thus 9001-2008 and API-Q1. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 6709 JOB 7305

Standard ID	0.100"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.098	Nominal	0.100	Maximum	0.102
Ref Type	5 Step Block	As Found	0.099	Accuracy	-0.001	Fail Before	No
		As Left	0.099	Accuracy	-0.001	Fail After	No
Gage ID of Standard	Ref-449809						
Std Due Date	4/10/2019	Gage S/N	449809	NIST No.	66030		

Standard ID	0.200"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.198	Nominal	0.200	Maximum	0.202
Ref Type	5 Step Block	As Found	0.200	Accuracy	0.000	Fail Before	No
		As Left	0.200	Accuracy	0.000	Fail After	No
Gage ID of Standard	Ref-449809						
Std Due Date	4/10/2019	Gage S/N	449809	NIST No.	66030		

Standard ID	0.300"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.298	Nominal	0.300	Maximum	0.302

Ref Type	5 Step Block	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard Ref-449809
 Std Due Date 4/10/2019

Gage S/N 449809 NIST No. 66030

Standard ID	0.400"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	5 Step Block	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard Ref-449809
 Std Due Date 4/10/2019

Gage S/N 449809 NIST No. 66030

Standard ID	0.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	5 Step Block	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard Ref-449809
 Std Due Date 4/10/2019

Gage S/N 449809 NIST No. 66030

Calibrated By Taichi Daimo Signature  Date: 10/16/2018